

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101718\
 Data File : VW006145.D
 Acq On : 18 Oct 2018 01:29
 Operator : VA/AP
 Sample : VW1017SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1017SBL02

Quant Time: Oct 18 13:45:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	267841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	394312	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	361735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	194392	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	112854	50.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.64%	
35) Dibromofluoromethane	7.89	113	111625	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
50) Toluene-d8	10.33	98	418825	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	
62) 4-Bromofluorobenzene	12.62	95	157136	47.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101718\
Data File : VW006145.D
Acq On : 18 Oct 2018 01:29
Operator : VA/AP
Sample : VW1017SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1017SBL02

Quant Time: Oct 18 13:45:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

