

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101718\
 Data File : VW006152.D
 Acq On : 18 Oct 2018 04:30
 Operator : VA/AP
 Sample : J5570-01
 Misc : 6.34G/5ML/MSVOA W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TS-06

Quant Time: Oct 18 14:09:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 16 05:36:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	241690	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	343609	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	274137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	110194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	98122	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	7.88	113	96375	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
50) Toluene-d8	10.33	98	299727	38.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.60%	
62) 4-Bromofluorobenzene	12.62	95	102279	35.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.50%	
Target Compounds						
20) Methylene Chloride	4.92	84	15038	5.649	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101718\
Data File : VW006152.D
Acq On : 18 Oct 2018 04:30
Operator : VA/AP
Sample : J5570-01
Misc : 6.34G/5ML/MSVOA W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TS-06

Quant Time: Oct 18 14:09:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 16 05:36:33 2018
Response via : Initial Calibration

