

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW101818\
 Data File : VW006218.D
 Acq On : 19 Oct 2018 11:40
 Operator : VA/AP
 Sample : J5590-01
 Misc : 6.39G/5ML/MSVOA_W/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 LS-SOIL-1

Manual Integrations
 APPROVED

apatel
 10/19/2018 3:19:37 PM

Quant Time: Oct 19 13:43:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	233368	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	345169	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	290848	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	123655	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	95424	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
35) Dibromofluoromethane	7.89	113	97243	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	10.33	98	344693	44.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.84%	
62) 4-Bromofluorobenzene	12.62	95	110923	38.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.12%	
Target Compounds						
16) Acetone	4.15	43	29907	71.85	ug/l	Qvalue 99
20) Methylene Chloride	4.92	84	25694	11.29	ug/l	96
25) 2-Butanone	7.19	43	4976	9.52	ug/l #	82
67) Ethyl Benzene	11.74	91	11796m	1.09	ug/l	
68) m/p-Xylenes	11.84	106	19681	4.61	ug/l	97
69) o-Xylene	12.17	106	6359	1.56	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW101818\
Data File : VW006218.D
Acq On : 19 Oct 2018 11:40
Operator : VA/AP
Sample : J5590-01
Misc : 6.39G/5ML/MSVOA_W/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
LS-SOIL-1

Manual Integrations
APPROVED

apatel
10/19/2018 3:19:37 PM

Quant Time: Oct 19 13:43:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

