

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
 Data File : VW020551.D
 Acq On : 18 Oct 2021 10:30
 Operator : SY/VA
 Sample : M4213-01RE
 Misc : 3.64g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGE3RE

Quant Time: Oct 19 05:42:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 19 05:41:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	222915	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	171009	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	44373	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	70380	21.698	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.800%
7) Chloroethane-d5	2.879	69	44407	25.918	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.680%
11) 1,1-Dichloroethene-d2	4.013	63	106359	16.002	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.000%
21) 2-Butanone-d5	7.080	46	32715	33.141	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.280%
24) Chloroform-d	7.647	84	163126	22.957	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.840%
26) 1,2-Dichloroethane-d4	8.305	65	88388	21.780	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.120%
32) Benzene-d6	8.275	84	302348	26.879	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.520%
36) 1,2-Dichloropropane-d6	9.275	67	99212	27.062	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.240%
41) Toluene-d8	10.323	98	243596	24.135	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.560%
43) trans-1,3-Dichloroprop...	10.579	79	31833	21.368	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.480%
47) 2-Hexanone-d5	10.927	63	22340	39.612	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.220%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	52297	18.872	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.480%
66) 1,2-Dichlorobenzene-d4	13.853	152	36873	21.243	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
Data File : VW020551.D
Acq On : 18 Oct 2021 10:30
Operator : SY/VA
Sample : M4213-01RE
Misc : 3.64g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGE3RE

Quant Time: Oct 19 05:42:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 19 05:41:45 2021
Response via : Initial Calibration

