

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
 Data File : VW020561.D
 Acq On : 18 Oct 2021 14:32
 Operator : SY/VA
 Sample : M4211-06
 Misc : 5.85g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGA5

Quant Time: Oct 19 05:43:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 19 05:41:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	235632	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	205813	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	85391	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	67630	19.725	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.880%
7) Chloroethane-d5	2.873	69	32220	17.790	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.160%
11) 1,1-Dichloroethene-d2	4.013	63	103093	14.674	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.680%
21) 2-Butanone-d5	7.080	46	48734	46.704	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.400%
24) Chloroform-d	7.647	84	161480	21.499	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	8.305	65	95556	22.275	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.120%
32) Benzene-d6	8.275	84	303159	22.393	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.560%
36) 1,2-Dichloropropane-d6	9.275	67	101891	23.093	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.360%
41) Toluene-d8	10.323	98	262855	21.639	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.560%
43) trans-1,3-Dichloroprop...	10.579	79	37964	21.174	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.680%
47) 2-Hexanone-d5	10.927	63	33308	49.073	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.140%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	73764	22.117	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.480%
66) 1,2-Dichlorobenzene-d4	13.853	152	77320	23.147	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.600%
Target Compounds					Qvalue	
13) Acetone	4.135	43	4670	4.875	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
Data File : VW020561.D
Acq On : 18 Oct 2021 14:32
Operator : SY/VA
Sample : M4211-06
Misc : 5.85g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGA5

Quant Time: Oct 19 05:43:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 19 05:41:45 2021
Response via : Initial Calibration

