

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
 Data File : VW020568.D
 Acq On : 18 Oct 2021 17:21
 Operator : SY/VA
 Sample : M4211-17
 Misc : 4.92g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGD7

Quant Time: Oct 19 05:44:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 19 05:41:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	242941	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	223346	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	103720	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	63250	17.892	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.560%
7) Chloroethane-d5	2.879	69	28713	15.377	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.520%
11) 1,1-Dichloroethene-d2	4.007	63	95680	13.209	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.840%
21) 2-Butanone-d5	7.080	46	45424	42.222	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.440%
24) Chloroform-d	7.647	84	160022	20.664	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.640%
26) 1,2-Dichloroethane-d4	8.305	65	92431	20.898	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.600%
32) Benzene-d6	8.275	84	298035	20.286	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.160%
36) 1,2-Dichloropropane-d6	9.274	67	100354	20.959	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.840%
41) Toluene-d8	10.323	98	259073	19.654	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.600%
43) trans-1,3-Dichloroprop...	10.579	79	37235	19.137	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.560%
47) 2-Hexanone-d5	10.927	63	29784	40.436	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.880%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	73076	20.191	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	87552	21.579	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.320%
Target Compounds						Qvalue
13) Acetone	4.135	43	8085	8.186	ug/L	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
Data File : VW020568.D
Acq On : 18 Oct 2021 17:21
Operator : SY/VA
Sample : M4211-17
Misc : 4.92g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGD7

Quant Time: Oct 19 05:44:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 19 05:41:45 2021
Response via : Initial Calibration

