

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
 Data File : VW020548.D
 Acq On : 18 Oct 2021 09:09
 Operator : SY/VA
 Sample : VW1018SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK464

Quant Time: Oct 19 05:42:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 19 05:41:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	181408	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	171041	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	80368	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	54230	20.544	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.160%
7) Chloroethane-d5	2.879	69	30065	21.562	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.240%
11) 1,1-Dichloroethene-d2	4.013	63	80496	14.882	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.520%
21) 2-Butanone-d5	7.080	46	31488	39.196	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.400%
24) Chloroform-d	7.653	84	121977	21.094	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.360%
26) 1,2-Dichloroethane-d4	8.311	65	69613	21.078	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.320%
32) Benzene-d6	8.275	84	230938	20.526	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.120%
36) 1,2-Dichloropropane-d6	9.274	67	77046	21.012	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.040%
41) Toluene-d8	10.323	98	204267	20.235	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.920%
43) trans-1,3-Dichloroprop...	10.579	79	29050	19.496	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.000%
47) 2-Hexanone-d5	10.927	63	20467	36.284	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.560%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	55333	19.964	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.840%
66) 1,2-Dichlorobenzene-d4	13.853	152	68815	21.889	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101821\
Data File : VW020548.D
Acq On : 18 Oct 2021 09:09
Operator : SY/VA
Sample : VW1018SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK464

Quant Time: Oct 19 05:42:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 19 05:41:45 2021
Response via : Initial Calibration

