

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101918\
 Data File : VW006226.D
 Acq On : 19 Oct 2018 15:46
 Operator : VA/AP
 Sample : J5575-03
 Misc : 5.77G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SSSI-1015-S-DH-27

Quant Time: Oct 20 01:55:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	258682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	373756	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	334994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	176668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	100876	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
35) Dibromofluoromethane	7.88	113	102614	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
50) Toluene-d8	10.33	98	385353	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
62) 4-Bromofluorobenzene	12.62	95	141345	44.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.58%	
Target Compounds						
16) Acetone	4.14	43	32621	70.445	ug/l	97
20) Methylene Chloride	4.92	84	6825	1.426	ug/l	87
52) Toluene	10.39	92	10501	1.658	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101918\
Data File : VW006226.D
Acq On : 19 Oct 2018 15:46
Operator : VA/AP
Sample : J5575-03
Misc : 5.77G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SSSI-1015-S-DH-27

Quant Time: Oct 20 01:55:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

