

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW101918\
 Data File : VW006242.D
 Acq On : 19 Oct 2018 23:05
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:51:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.247	0.252	-2.0	98	0.00
3 P	Chloromethane	0.318	0.319	-0.3	105	0.00
4 C	Vinyl Chloride	0.455	0.473	-4.0#	104	0.00
5 T	Bromomethane	0.414	0.380	8.2	96	0.00
6 T	Chloroethane	0.326	0.325	0.3	100	0.00
7 T	Trichlorofluoromethane	0.424	0.384	9.4	89	0.00
8 T	Diethyl Ether	0.225	0.241	-7.1	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.444	0.462	-4.1	108	0.00
10 T	Methyl Iodide	0.739	0.811	-9.7	112	0.00
11 T	Tert butyl alcohol	0.028	0.028	0.0	112	0.03
12 CM	1,1-Dichloroethene	0.418	0.441	-5.5#	110	0.00
13 T	Acrolein	0.027	0.027	0.0	109	0.00
14 T	Allyl chloride	0.597	0.625	-4.7	108	0.00
15 T	Acrylonitrile	0.081	0.092	-13.6	118	0.00
16 T	Acetone	0.089	0.078	12.4	110	0.00
17 T	Carbon Disulfide	1.261	1.209	4.1	98	0.00
18 T	Methyl Acetate	0.209	0.230	-10.0	114	0.00
19 T	Methyl tert-butyl Ether	0.636	0.710	-11.6	116	0.00
20 T	Methylene Chloride	0.484	0.488	-0.8	115	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.508	-9.0	112	0.00
22 T	Diisopropyl ether	1.180	1.296	-9.8	113	0.00
23 T	Vinyl Acetate	0.697	0.757	-8.6	113	0.00
24 P	1,1-Dichloroethane	0.762	0.842	-10.5	113	0.00
25 T	2-Butanone	0.112	0.118	-5.4	115	0.00
26 T	2,2-Dichloropropane	0.537	0.493	8.2	99	0.00
27 T	cis-1,2-Dichloroethene	0.515	0.574	-11.5	115	0.00
28 T	Bromochloromethane	0.294	0.329	-11.9	108	0.00
29 T	Tetrahydrofuran	0.069	0.076	-10.1	117	0.00
30 C	Chloroform	0.830	0.925	-11.4#	116	0.00
31 T	Cyclohexane	0.759	0.719	5.3	104	0.00
32 T	1,1,1-Trichloroethane	0.749	0.776	-3.6	107	0.00
33 S	1,2-Dichloroethane-d4	0.415	0.418	-0.7	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.295	0.294	0.3	102	0.00
36 T	1,1-Dichloropropene	0.461	0.474	-2.8	110	0.00
37 T	Ethyl Acetate	0.171	0.184	-7.6	120	0.00
38 T	Carbon Tetrachloride	0.505	0.496	1.8	104	0.00
39 T	Methylcyclohexane	0.613	0.612	0.2	107	0.00
40 TM	Benzene	1.264	1.353	-7.0	114	0.00
41 T	Methacrylonitrile	0.103	0.110	-6.8	112	0.00
42 TM	1,2-Dichloroethane	0.371	0.410	-10.5	117	0.00
43 T	Isopropyl Acetate	0.350	0.370	-5.7	114	0.00
44 TM	Trichloroethene	0.388	0.415	-7.0	114	0.00
45 C	1,2-Dichloropropane	0.303	0.323	-6.6#	114	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW101918\
 Data File : VW006242.D
 Acq On : 19 Oct 2018 23:05
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:51:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.172	0.189	-9.9	116	0.00
47 T	Bromodichloromethane	0.418	0.439	-5.0	110	0.00
48 T	Methyl methacrylate	0.166	0.181	-9.0	115	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	121	0.02
50 S	Toluene-d8	1.124	1.076	4.3	99	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.189	-8.0	116	0.00
52 CM	Toluene	0.848	0.895	-5.5#	113	0.00
53 T	t-1,3-Dichloropropene	0.416	0.434	-4.3	109	0.00
54 T	cis-1,3-Dichloropropene	0.478	0.499	-4.4	110	0.00
55 T	1,1,2-Trichloroethane	0.241	0.268	-11.2	118	0.00
56 T	Ethyl methacrylate	0.296	0.332	-12.2	119	0.00
57 T	1,3-Dichloropropane	0.399	0.439	-10.0	117	0.00
58 T	2-Chloroethyl Vinyl ether	0.113	0.114	-0.9	111	0.00
59 T	2-Hexanone	0.125	0.129	-3.2	113	0.00
60 T	Dibromochloromethane	0.301	0.320	-6.3	110	0.00
61 T	1,2-Dibromoethane	0.241	0.271	-12.4	118	0.00
62 S	4-Bromofluorobenzene	0.422	0.426	-0.9	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.388	0.407	-4.9	113	0.00
65 PM	Chlorobenzene	1.070	1.141	-6.6	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.404	-6.3	111	0.00
67 C	Ethyl Benzene	1.855	1.947	-5.0#	112	0.00
68 T	m/p-Xylenes	0.733	0.774	-5.6	112	0.00
69 T	o-Xylene	0.699	0.743	-6.3	113	0.00
70 T	Styrene	1.136	1.230	-8.3	114	0.00
71 P	Bromoform	0.212	0.216	-1.9	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.532	3.674	-4.0	111	0.00
74 T	N-amyl acetate	0.686	0.719	-4.8	112	0.00
75 P	1,1,2,2-Tetrachloroethane	0.537	0.598	-11.4	117	0.00
76 T	1,2,3-Trichloropropane	0.382	0.390	-2.1	113	0.00
77 T	Bromobenzene	0.865	0.938	-8.4	115	0.00
78 T	n-propylbenzene	4.110	4.256	-3.6	110	0.00
79 T	2-Chlorotoluene	2.309	2.434	-5.4	111	0.00
80 T	1,3,5-Trimethylbenzene	2.978	3.112	-4.5	110	0.00
81 T	trans-1,4-Dichloro-2-butene	0.160	0.154	3.8	100	0.00
82 T	4-Chlorotoluene	2.419	2.578	-6.6	113	0.00
83 T	tert-Butylbenzene	2.664	2.769	-3.9	110	0.00
84 T	1,2,4-Trimethylbenzene	3.013	3.189	-5.8	112	0.00
85 T	sec-Butylbenzene	3.764	3.889	-3.3	109	0.00
86 T	p-Isopropyltoluene	3.404	3.552	-4.3	109	0.00
87 T	1,3-Dichlorobenzene	1.722	1.841	-6.9	114	0.00
88 T	1,4-Dichlorobenzene	1.714	1.813	-5.8	113	0.00
89 T	n-Butylbenzene	3.114	3.227	-3.6	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW101918\
Data File : VW006242.D
Acq On : 19 Oct 2018 23:05
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 20 03:51:21 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.562	2.6	99	0.00
91 T	1,2-Dichlorobenzene	1.536	1.651	-7.5	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.098	-2.1	108	0.00
93 T	1,2,4-Trichlorobenzene	1.182	1.268	-7.3	117	0.00
94 T	Hexachlorobutadiene	0.742	0.762	-2.7	110	0.00
95 T	Naphthalene	1.966	2.241	-14.0	123	0.00
96 T	1,2,3-Trichlorobenzene	1.022	1.124	-10.0	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6