

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
 Data File : VW020608.D
 Acq On : 19 Oct 2021 14:43
 Operator : SY/VA
 Sample : M4207-11
 Misc : 4.34g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B33

Quant Time: Oct 20 01:39:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 20 01:35:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	205203	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	180376	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	68201	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	56562	18.943	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.760%
7) Chloroethane-d5	2.879	69	26348	16.705	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.840%
11) 1,1-Dichloroethene-d2	4.013	63	87240	14.259	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.040%
21) 2-Butanone-d5	7.080	46	41542	45.715	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.440%
24) Chloroform-d	7.647	84	155580	23.785	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.120%
26) 1,2-Dichloroethane-d4	8.305	65	92773	24.833	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.320%
32) Benzene-d6	8.275	84	276611	23.314	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.240%
36) 1,2-Dichloropropane-d6	9.275	67	98122	25.375	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.480%
41) Toluene-d8	10.323	98	232006	21.793	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.160%
43) trans-1,3-Dichloroprop...	10.579	79	33560	21.357	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.440%
47) 2-Hexanone-d5	10.927	63	27191	45.710	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.420%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	75232	25.739	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.960%
66) 1,2-Dichlorobenzene-d4	13.853	152	66611	24.968	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
Data File : VW020608.D
Acq On : 19 Oct 2021 14:43
Operator : SY/VA
Sample : M4207-11
Misc : 4.34g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B33

Quant Time: Oct 20 01:39:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 20 01:35:36 2021
Response via : Initial Calibration

