

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
 Data File : VW020609.D
 Acq On : 19 Oct 2021 15:07
 Operator : SY/VA
 Sample : M4212-05
 Misc : 4.70g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JDGB8

Quant Time: Oct 20 01:39:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 20 01:35:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	220671	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	185332	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	66631	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.348	65	61346	19.105	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.440%
7) Chloroethane-d5	2.885	69	27156	16.010	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.040%
11) 1,1-Dichloroethene-d2	4.013	63	79997	12.158	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.640%
21) 2-Butanone-d5	7.079	46	38291	39.184	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.360%
24) Chloroform-d	7.652	84	157390	22.375	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.520%
26) 1,2-Dichloroethane-d4	8.305	65	86890	21.628	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.520%
32) Benzene-d6	8.274	84	281145	23.062	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.240%
36) 1,2-Dichloropropane-d6	9.274	67	95953	24.150	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.600%
41) Toluene-d8	10.323	98	229632	20.993	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.960%
43) trans-1,3-Dichloroprop...	10.579	79	32512	20.137	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.560%
47) 2-Hexanone-d5	10.926	63	23553	38.535	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.080%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	63463	21.131	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.520%
66) 1,2-Dichlorobenzene-d4	13.853	152	57540	22.076	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
Data File : VW020609.D
Acq On : 19 Oct 2021 15:07
Operator : SY/VA
Sample : M4212-05
Misc : 4.70g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JDGB8

Quant Time: Oct 20 01:39:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 20 01:35:36 2021
Response via : Initial Calibration

