

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
 Data File : VW020613.D
 Acq On : 19 Oct 2021 16:43
 Operator : SY/VA
 Sample : M4263-01
 Misc : 7.57g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB7E6

Quant Time: Oct 20 01:40:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 20 01:35:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	228304	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	213851	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	93939	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	65076	19.589	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.360%
7) Chloroethane-d5	2.879	69	27823	15.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.440%
11) 1,1-Dichloroethene-d2	4.013	63	100611	14.780	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.120%
21) 2-Butanone-d5	7.080	46	49930	49.386	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.780%
24) Chloroform-d	7.647	84	152289	20.926	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.720%
26) 1,2-Dichloroethane-d4	8.305	65	95034	22.864	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.440%
32) Benzene-d6	8.275	84	301623	21.442	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.760%
36) 1,2-Dichloropropane-d6	9.274	67	100367	21.892	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.560%
41) Toluene-d8	10.323	98	276126	21.877	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.520%
43) trans-1,3-Dichloroprop...	10.579	79	41913	22.497	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.000%
47) 2-Hexanone-d5	10.927	63	37385	53.009	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.020%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	78618	22.687	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	85258	23.201	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.800%
Target Compounds					Qvalue	
13) Acetone	4.129	43	22294	24.018	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
Data File : VW020613.D
Acq On : 19 Oct 2021 16:43
Operator : SY/VA
Sample : M4263-01
Misc : 7.57g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GB7E6

Quant Time: Oct 20 01:40:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 20 01:35:36 2021
Response via : Initial Calibration

