

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
 Data File : VW020606.D
 Acq On : 19 Oct 2021 13:55
 Operator : SY/VA
 Sample : M4207-09
 Misc : 5.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B22

Quant Time: Oct 20 01:39:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 20 01:35:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	165935	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	119878	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	26324	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	60693	25.137	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.560%	
7) Chloroethane-d5	2.879	69	32132	25.193	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.760%	
11) 1,1-Dichloroethene-d2	4.013	63	90581	18.308	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 73.240%	
21) 2-Butanone-d5	7.086	46	38948	53.003	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 106.000%	
24) Chloroform-d	7.653	84	161591	30.550	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 122.200%	
26) 1,2-Dichloroethane-d4	8.305	65	88886	29.423	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 117.680%	
32) Benzene-d6	8.274	84	274289	34.785	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 139.120%#	
36) 1,2-Dichloropropane-d6	9.274	67	100792	39.219	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 156.880%#	
41) Toluene-d8	10.323	98	212794	30.076	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 120.320%	
43) trans-1,3-Dichloroprop...	10.579	79	26867	25.726	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 102.920%	
47) 2-Hexanone-d5	10.926	63	24338	61.561	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 123.120%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	66034	33.993	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 135.960%#	
66) 1,2-Dichlorobenzene-d4	13.853	152	33240	32.280	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 129.120%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101921\
Data File : VW020606.D
Acq On : 19 Oct 2021 13:55
Operator : SY/VA
Sample : M4207-09
Misc : 5.16g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B22

Quant Time: Oct 20 01:39:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 20 01:35:36 2021
Response via : Initial Calibration

