

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\WV102119\
 Data File : VW013631.D
 Acq On : 21 Oct 2019 13:07
 Operator : SY/VA
 Sample : K5395-14
 Misc : 4.56G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0030

Quant Time: Oct 21 17:16:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 21 16:37:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	343265	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	298395	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	131247	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	118586	28.16	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.64%
7) Chloroethane-d5	2.89	69	100719	25.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.48%
10) 1,1-Dichloroethene-d2	4.02	63	179930	19.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.36%
20) 2-Butanone-d5	7.08	46	59648	39.60	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.20%
24) Chloroform-d	7.65	84	226497	23.58	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.32%
26) 1,2-Dichloroethane-d4	8.30	65	122103	24.31	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.24%
29) Benzene-d6	8.27	84	502650	28.04	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	112.16%
33) 1,2-Dichloropropane-d6	9.27	67	146599	27.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	109.52%
37) Toluene-d8	10.32	98	447495	27.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	108.24%
38) trans-1,3-Dichloropropene-	10.58	79	58670	24.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.16%
39) 2-Hexanone-d5	10.92	63	46160	45.10	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.20%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	99971	22.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.20%
61) 1,2-Dichlorobenzene-d4	13.85	152	114971	24.79	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.16%


Target Compounds					Ovalue
13) Acetone	4.13	43	8998	5.382 ug/L	74
14) Carbon disulfide	4.39	76	37689	2.414 ug/L	98
16) Methylene chloride	4.92	84	55587	8.457 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102119\
Data File : VW013631.D
Acq On : 21 Oct 2019 13:07
Operator : SY/VA
Sample : K5395-14
Misc : 4.56G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0030

Quant Time: Oct 21 17:16:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 21 16:37:58 2019
Response via : Initial Calibration





#13

Acetone

Concen: 5.382 ug/L

RT: 4.13 min Scan# 366

Delta R.T. 0.01 min

Lab File: VW013631.D

Acq: 21 Oct 2019 13:07

Instrument :

MSVOA_W

ClientSampleId :


C0030

Tot Ion: 43 Resp: 8998

Ion Ratio Lower Upper




43 100

58 19.8 0.0 70.4



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW



#14

Carbon disulfide

Concen: 2.414 ug/L

RT: 4.39 min Scan# 408

Delta R.T. 0.01 min

Lab File: VW013631.D


Acq: 21 Oct 2019 13:07

Tgt Ion: 76 Resp: 37689

Ion Ratio Lower Upper



76 100


78 8.0 7.1 10.7



Abundance Ion 76.00 (75.70 to 76.70): VW

Ion 78.00 (77.70 to 78.70): VW





#16
 Methylene chloride
 Concen: 8.457 ug/L
 RT: 4.92 min Scan# 495
 Delta R.T. -0.00 min
 Lab File: VW013631.D
 Acq: 21 Oct 2019 13:07

Instrument :
 MSVOA_W
 ClientSampleId :
 C0030

Tot Ion: 84 Resp: 55587
 Ion Ratio Lower Upper
 84 100
 49 116.2 80.2 148.9
 86 60.0 42.7 79.3

