

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102120\
 Data File : VW016957.D
 Acq On : 21 Oct 2020 15:10
 Operator : SY/VA
 Sample : VIBLK21
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK21

Quant Time: Oct 22 12:56:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 21 22:50:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	475638	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	429670	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	205659	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	80031	17.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.24%
7) Chloroethane-d5	2.89	69	73100	18.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.28%
10) 1,1-Dichloroethene-d2	4.03	63	135862	13.64	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.56%
20) 2-Butanone-d5	7.09	46	59525	56.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.02%
24) Chloroform-d	7.65	84	228247	22.74	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.96%
26) 1,2-Dichloroethane-d4	8.31	65	113219	23.36	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.44%
29) Benzene-d6	8.27	84	435764	22.33	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.32%
33) 1,2-Dichloropropane-d6	9.28	67	129988	24.79	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.16%
37) Toluene-d8	10.33	98	400713	21.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.04%
38) trans-1,3-Dichloropropene-	10.58	79	50761	21.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.80%
39) 2-Hexanone-d5	10.93	63	48275	59.52	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.04%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	121583	28.70	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	114.80%
61) 1,2-Dichlorobenzene-d4	13.85	152	163149	25.01	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.04%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	6299	0.960 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102120\
Data File : VW016957.D
Acq On : 21 Oct 2020 15:10
Operator : SY/VA
Sample : VIBLK21
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK21

Quant Time: Oct 22 12:56:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100720S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 21 22:50:53 2020
Response via : Initial Calibration

