

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102218\
 Data File : VW006276.D
 Acq On : 23 Oct 2018 04:03
 Operator : VA/AP
 Sample : J5625-16
 Misc : 4.55G/5ML/MSVOA W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-08

Quant Time: Oct 23 06:56:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	226600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	306972	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	174833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	36375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	60129	32.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	64.02%	
35) Dibromofluoromethane	7.88	113	64777	35.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.42%	
50) Toluene-d8	10.33	98	179409	26.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	52.00%	
62) 4-Bromofluorobenzene	12.62	95	33556	12.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	25.90%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.91	84	10470	3.762	ug/l	94
67) Ethyl Benzene	11.74	91	43400	6.691	ug/l	99
68) m/p-Xylenes	11.84	106	65678	25.619	ug/l	98
69) o-Xylene	12.17	106	21660	8.865	ug/l	97
95) Naphthalene	15.38	128	3493	2.442	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102218\
Data File : VW006276.D
Acq On : 23 Oct 2018 04:03
Operator : VA/AP
Sample : J5625-16
Misc : 4.55G/5ML/MSVOA W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-08

Quant Time: Oct 23 06:56:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

