

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102218\
 Data File : VW006284.D
 Acq On : 23 Oct 2018 07:30
 Operator : VA/AP
 Sample : J5621-07
 Misc : 4.65G/5ML/MSVOA W/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SA-SB-01C

Quant Time: Oct 23 08:38:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	211232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	307978	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	243872	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	100915	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	60094	34.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.64%	
35) Dibromofluoromethane	7.88	113	62289	34.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.46%	
50) Toluene-d8	10.33	98	193013	27.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.76%	
62) 4-Bromofluorobenzene	12.62	95	58574	22.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	45.04%	
Target Compounds						
20) Methylene Chloride	4.91	84	9078	3.382	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102218\
Data File : VW006284.D
Acq On : 23 Oct 2018 07:30
Operator : VA/AP
Sample : J5621-07
Misc : 4.65G/5ML/MSVOA W/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SA-SB-01C

Quant Time: Oct 23 08:38:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

