

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102218\
 Data File : VW006290.D
 Acq On : 23 Oct 2018 10:05
 Operator : VA/AP
 Sample : J5624-06
 Misc : 4.41G/5ML/MSVOA W/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PA-SB-03B

Quant Time: Oct 23 15:07:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	242558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	363244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	308222	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	134764	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	68447	34.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.08%	
35) Dibromofluoromethane	7.89	113	68632	31.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.96%	
50) Toluene-d8	10.33	98	221886	27.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.34%	
62) 4-Bromofluorobenzene	12.62	95	75159	24.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.02%	
Target Compounds						
16) Acetone	4.13	43	6978	3.925	ug/l	93
17) Carbon Disulfide	4.37	76	13249	2.166	ug/l	98
20) Methylene Chloride	4.91	84	8447	2.421	ug/l	91
27) cis-1,2-Dichloroethene	7.17	96	2534	1.015	ug/l	95
44) Trichloroethene	9.09	130	10238	3.634	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102218\
Data File : VW006290.D
Acq On : 23 Oct 2018 10:05
Operator : VA/AP
Sample : J5624-06
Misc : 4.41G/5ML/MSVOA W/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PA-SB-03B

Quant Time: Oct 23 15:07:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

