

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\WV102220\
 Data File : VW016971.D
 Acq On : 22 Oct 2020 15:48
 Operator : SY/VA
 Sample : L4471-17
 Misc : 5.79G/10ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG432

Quant Time: Oct 23 02:48:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 22 12:40:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	611282	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	550722	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	256411	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	164786	23.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.44%
7) Chloroethane-d5	2.89	69	124523	22.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.12%
10) 1,1-Dichloroethene-d2	4.02	63	232349	16.93	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.72%
20) 2-Butanone-d5	7.09	46	69257	52.11	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.22%
24) Chloroform-d	7.65	84	314060	22.27	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.08%
26) 1,2-Dichloroethane-d4	8.31	65	161875	24.39	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.56%
29) Benzene-d6	8.28	84	657768	23.25	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.00%
33) 1,2-Dichloropropane-d6	9.27	67	181181	23.73	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.92%
37) Toluene-d8	10.32	98	617995	23.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.80%
38) trans-1,3-Dichloropropene-	10.58	79	76324	23.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.20%
39) 2-Hexanone-d5	10.93	63	59013	55.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.08%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	147458	25.35	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.40%
61) 1,2-Dichlorobenzene-d4	13.85	152	218784	24.28	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.12%

Target Compounds					Ovalue
13) Acetone	4.14	43	6297	7.042 ug/L	96
14) Carbon disulfide	4.39	76	7977	0.329 ug/L #	92
16) Methylene chloride	4.92	84	34570	4.277 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102220\
Data File : VW016971.D
Acq On : 22 Oct 2020 15:48
Operator : SY/VA
Sample : L4471-17
Misc : 5.79G/10ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG432

Quant Time: Oct 23 02:48:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 12:40:20 2020
Response via : Initial Calibration

