


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102319\
Data File : VW013660.D
Acq On : 23 Oct 2019 10:20
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
Title : SW846 8260
Last Update : Wed Oct 23 14:11:46 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	24485	PASS
75	95	30	60	47.2	72170	PASS
95	95	100	100	100.0	152944	PASS
96	95	5	9	6.5	9985	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	93.2	142498	PASS
175	174	5	9	8.1	11534	PASS
176	174	95	101	97.2	138530	PASS
177	176	5	9	6.8	9460	PASS