

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102419\
 Data File : VW013711.D
 Acq On : 24 Oct 2019 15:16
 Operator : SY/VA
 Sample : VW1024SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1024SBL01

Quant Time: Oct 25 08:06:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 23 14:11:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	329566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	472347	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	408821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	198201	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	126457	51.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.30%	
35) Dibromofluoromethane	7.88	113	131524	50.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.22%	
50) Toluene-d8	10.32	98	530873	49.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.62	95	173302	47.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.72%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102419\
Data File : VW013711.D
Acq On : 24 Oct 2019 15:16
Operator : SY/VA
Sample : VW1024SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1024SBL01

Quant Time: Oct 25 08:06:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 23 14:11:46 2019
Response via : Initial Calibration

