


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102419\
Data File : VW013709.D
Acq On : 24 Oct 2019 13:28
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
Title : SW846 8260
Last Update : Wed Oct 23 14:11:46 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	40000	PASS
75	95	30	60	47.5	113837	PASS
95	95	100	100	100.0	239722	PASS
96	95	5	9	6.6	15853	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.2	213866	PASS
175	174	5	9	8.0	17039	PASS
176	174	95	101	96.0	205418	PASS
177	176	5	9	6.7	13723	PASS