

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102424\
 Data File : VW030739.D
 Acq On : 24 Oct 2024 12:31
 Operator : SY/MD
 Sample : VW1024SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK581

Quant Time: Oct 25 01:43:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 25 01:42:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	268481	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	256812	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	113086	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	52381	12.412	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	49.640%
7) Chloroethane-d5	2.899	69	51191	16.025	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.080%
11) 1,1-Dichloroethene-d2	4.021	65	23199	13.217	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.880%
21) 2-Butanone-d5	7.081	46	34433	45.339	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.680%
24) Chloroform-d	7.654	84	141407	18.417	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.680%
26) 1,2-Dichloroethane-d4	8.307	65	77117	19.315	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.280%
32) Benzene-d6	8.276	84	239814	16.089	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.360%
36) 1,2-Dichloropropane-d6	9.276	67	80139	18.318	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	73.280%
41) Toluene-d8	10.325	98	205479	15.101	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	60.400%
43) trans-1,3-Dichloroprop...	10.581	79	28795	15.719	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	62.880%
47) 2-Hexanone-d5	10.922	63	23878	42.376	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.760%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	70756	22.023	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.080%
66) 1,2-Dichlorobenzene-d4	13.854	152	78086	20.407	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102424\
Data File : VW030739.D
Acq On : 24 Oct 2024 12:31
Operator : SY/MD
Sample : VW1024SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK581

Quant Time: Oct 25 01:43:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 25 01:42:56 2024
Response via : Initial Calibration

