

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102519\
 Data File : VW013739.D
 Acq On : 25 Oct 2019 11:19
 Operator : SY/VA
 Sample : K5535-05
 Misc : 6.33G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BEK27

Manual Integrations
 APPROVED

MMDadoda
 10/29/2019 2:51:47 PM

Quant Time: Oct 26 03:05:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 26 03:37:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	361141	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	275147	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	93476	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	131872	29.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.04%
7) Chloroethane-d5	2.89	69	107681	26.04	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.16%
10) 1,1-Dichloroethene-d2	4.02	63	191979	19.62	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.48%
20) 2-Butanone-d5	7.08	46	59960	37.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.68%
24) Chloroform-d	7.65	84	251355	24.87	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.48%
26) 1,2-Dichloroethane-d4	8.30	65	125130	23.68	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.72%
29) Benzene-d6	8.27	84	544536	32.94	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	131.76%
33) 1,2-Dichloropropane-d6	9.27	67	159217	32.25	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	129.00%#
37) Toluene-d8	10.32	98	472576	31.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	124.00%
38) trans-1,3-Dichloropropene-	10.58	79	53705	24.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.44%
39) 2-Hexanone-d5	10.93	63	44753	47.42	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.84%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	95984	23.74	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.96%
61) 1,2-Dichlorobenzene-d4	13.85	152	87893	26.61	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.44%

Target Compounds					Ovalue
13) Acetone	4.13	43	4914	2.794 ug/L	57
30) Cyclohexane	7.95	56	15691m	1.947 ug/L	
36) Methylcyclohexane	9.34	83	123475	13.830 ug/L #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102519\
Data File : VW013739.D
Acq On : 25 Oct 2019 11:19
Operator : SY/VA
Sample : K5535-05
Misc : 6.33G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BEK27

Manual Integrations
APPROVED

MMDadoda
10/29/2019 2:51:47 PM

Quant Time: Oct 26 03:05:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 26 03:37:49 2019
Response via : Initial Calibration

