

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102519\
 Data File : VW013755.D
 Acq On : 25 Oct 2019 18:17
 Operator : SY/VA
 Sample : K5549-12
 Misc : 4.13G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAQ65

Manual Integrations
 APPROVED

MMDadoda
 10/28/2019 9:41:39 AM

Quant Time: Oct 26 05:51:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 26 03:37:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	282511	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	280904	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	141316	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	88194	25.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.76%
7) Chloroethane-d5	2.89	69	71875	22.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.84%
10) 1,1-Dichloroethene-d2	4.03	63	132669	17.33	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.32%
20) 2-Butanone-d5	7.08	46	71387	57.58	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.16%
24) Chloroform-d	7.65	84	180114	22.78	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.12%
26) 1,2-Dichloroethane-d4	8.31	65	107639	26.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.16%
29) Benzene-d6	8.27	84	384644	22.79	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.16%
33) 1,2-Dichloropropane-d6	9.27	67	117653	23.34	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.36%
37) Toluene-d8	10.32	98	352413	22.64	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.56%
38) trans-1,3-Dichloropropene-	10.58	79	52430	23.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.24%
39) 2-Hexanone-d5	10.92	63	59712	61.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.94%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	112218	27.18	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.72%
61) 1,2-Dichlorobenzene-d4	13.85	152	119161	23.86	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.44%

Target Compounds

					Ovalue
13) Acetone	4.13	43	14165	10.295	ug/L 97
15) Methyl Acetate	4.69	43	3225m	1.434	ug/L
44) Toluene	10.38	91	19108	0.882	ug/L 93
47) Tetrachloroethene	10.87	164	3023	0.639	ug/L 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102519\
Data File : VW013755.D
Acq On : 25 Oct 2019 18:17
Operator : SY/VA
Sample : K5549-12
Misc : 4.13G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAQ65

Manual Integrations
APPROVED

MMDadoda
10/28/2019 9:41:39 AM

Quant Time: Oct 26 05:51:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 26 03:37:49 2019
Response via : Initial Calibration

