

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102622\
 Data File : VW024986.D
 Acq On : 26 Oct 2022 14:46
 Operator : SY/VA
 Sample : N5258-07RE
 Misc : 6.43g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBLN6RE

Quant Time: Oct 27 04:49:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 27 04:48:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	233383	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	112371	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	20886	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	57054	21.780	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.120%
7) Chloroethane-d5	2.885	69	36242	23.618	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.480%
11) 1,1-Dichloroethene-d2	4.019	63	92125	14.118	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.480%
21) 2-Butanone-d5	7.074	46	112561	127.317	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	254.640%#
24) Chloroform-d	7.653	84	177736	25.094	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.360%
26) 1,2-Dichloroethane-d4	8.305	65	123509m	31.152	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	124.600%
32) Benzene-d6	8.275	84	304842	42.324	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	169.280%#
36) 1,2-Dichloropropane-d6	9.274	67	100885	47.804	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	191.200%#
41) Toluene-d8	10.323	98	171508	25.577	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.320%
43) trans-1,3-Dichloroprop...	10.579	79	33800	36.119	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	144.480%#
47) 2-Hexanone-d5	10.920	63	75596	207.366	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	414.740%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	98856	58.666	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	234.680%#
66) 1,2-Dichlorobenzene-d4	13.853	152	18890	23.535	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.120%
Target Compounds						
					Qvalue	
13) Acetone	4.129	43	11286	17.969	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102622\
Data File : VW024986.D
Acq On : 26 Oct 2022 14:46
Operator : SY/VA
Sample : N5258-07RE
Misc : 6.43g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GBLN6RE

Quant Time: Oct 27 04:49:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM102622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 27 04:48:47 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 17.969 ug/L
 RT: 4.129 min Scan# 30
 Delta R.T. 0.006 min
 Lab File: VW024986.D
 Acq: 26 Oct 2022 14:46

Instrument :
 MSVOA_W
 ClientSampleId :
 GBLN6RE

Tgt Ion: 43 Resp: 11286
 Ion Ratio Lower Upper
 43 100
 58 30.8 0.0 65.4

