

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102818\
 Data File : VW006402.D
 Acq On : 26 Oct 2018 12:17
 Operator : SY/AP
 Sample : VW1028SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK94

Quant Time: Oct 27 00:04:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100918S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 26 23:43:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	331791	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	314553	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	167376	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	109517	29.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	117.08%
7) Chloroethane-d5	2.90	69	95487	29.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	118.60%
10) 1,1-Dichloroethene-d2	4.01	63	144937	20.17	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.68%
20) 2-Butanone-d5	7.09	46	59550	48.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.02%
24) Chloroform-d	7.65	84	190320	26.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.92%
26) 1,2-Dichloroethane-d4	8.31	65	111784	28.11	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.44%
29) Benzene-d6	8.27	84	398205	27.88	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	111.52%
33) 1,2-Dichloropropane-d6	9.27	67	111692	25.91	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.64%
37) Toluene-d8	10.32	98	393737	27.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	110.28%
38) trans-1,3-Dichloropropene-	10.58	79	47973	22.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.60%
39) 2-Hexanone-d5	10.93	63	49532	47.56	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.12%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	109206	26.06	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.24%
61) 1,2-Dichlorobenzene-d4	13.86	152	160406	27.96	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	111.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102818\
Data File : VW006402.D
Acq On : 26 Oct 2018 12:17
Operator : SY/AP
Sample : VW1028SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK94

Quant Time: Oct 27 00:04:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM100918S.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 26 23:43:44 2018
Response via : Initial Calibration

