

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102819\
 Data File : VW013759.D
 Acq On : 28 Oct 2019 11:43
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 10:59:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 23 14:11:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.263	0.225	14.4	100	0.00
3 P	Chloromethane	0.283	0.263	7.1	106	0.00
4 C	Vinyl Chloride	0.428	0.412	3.7#	100	0.00
5 T	Bromomethane	0.281	0.267	5.0	100	0.00
6 T	Chloroethane	0.259	0.248	4.2	97	0.00
7 T	Trichlorofluoromethane	0.267	0.278	-4.1	107	0.00
8 T	Diethyl Ether	0.216	0.198	8.3	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.434	0.408	6.0	95	0.00
10 T	Methyl Iodide	0.707	0.679	4.0	96	0.00
11 T	Tert butyl alcohol	0.022	0.020	9.1	96	0.00
12 CM	1,1-Dichloroethene	0.448	0.437	2.5#	98	0.00
13 T	Acrolein	0.026	0.032	-23.1#	125	0.00
14 T	Allyl chloride	0.690	0.678	1.7	98	0.00
15 T	Acrylonitrile	0.088	0.078	11.4	93	0.00
16 T	Acetone	0.080	0.084	-5.0	101	0.00
17 T	Carbon Disulfide	1.288	1.263	1.9	98	0.00
18 T	Methyl Acetate	0.218	0.188	13.8	94	0.00
19 T	Methyl tert-butyl Ether	0.538	0.528	1.9	98	0.00
20 T	Methylene Chloride	0.501	0.431	14.0	96	0.00
21 T	trans-1,2-Dichloroethene	0.483	0.470	2.7	97	0.00
22 T	Diisopropyl ether	1.258	1.223	2.8	96	0.00
23 T	Vinyl Acetate	0.730	0.696	4.7	94	0.00
24 P	1,1-Dichloroethane	0.797	0.766	3.9	96	0.00
25 T	2-Butanone	0.114	0.105	7.9	93	0.00
26 T	2,2-Dichloropropane	0.461	0.452	2.0	103	0.00
27 T	cis-1,2-Dichloroethene	0.513	0.492	4.1	95	0.00
28 T	Bromochloromethane	0.265	0.289	-9.1	93	0.00
29 T	Tetrahydrofuran	0.071	0.063	11.3	92	0.00
30 C	Chloroform	0.787	0.751	4.6#	97	0.00
31 T	Cyclohexane	0.834	0.777	6.8	97	0.00
32 T	1,1,1-Trichloroethane	0.618	0.615	0.5	100	0.00
33 S	1,2-Dichloroethane-d4	0.375	0.371	1.1	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.278	0.284	-2.2	94	0.00
36 T	1,1-Dichloropropene	0.466	0.462	0.9	98	0.00
37 T	Ethyl Acetate	0.178	0.157	11.8	93	0.00
38 T	Carbon Tetrachloride	0.419	0.421	-0.5	100	0.00
39 T	Methylcyclohexane	0.601	0.603	-0.3	97	0.00
40 TM	Benzene	1.318	1.288	2.3	97	0.00
41 T	Methacrylonitrile	0.111	0.104	6.3	101	0.00
42 TM	1,2-Dichloroethane	0.332	0.319	3.9	95	0.00
43 T	Isopropyl Acetate	0.341	0.315	7.6	93	0.00
44 TM	Trichloroethene	0.377	0.377	0.0	98	0.00
45 C	1,2-Dichloropropane	0.315	0.305	3.2#	95	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102819\
 Data File : VW013759.D
 Acq On : 28 Oct 2019 11:43
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 10:59:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 23 14:11:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.153	0.142	7.2	93	0.00
47 T	Bromodichloromethane	0.395	0.388	1.8	96	0.00
48 T	Methyl methacrylate	0.169	0.148	12.4	86	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	96	0.00
50 S	Toluene-d8	1.132	1.189	-5.0	95	0.00
51 T	4-Methyl-2-Pentanone	0.170	0.153	10.0	92	0.00
52 CM	Toluene	0.850	0.835	1.8#	96	0.00
53 T	t-1,3-Dichloropropene	0.400	0.396	1.0	96	0.00
54 T	cis-1,3-Dichloropropene	0.494	0.492	0.4	97	0.00
55 T	1,1,2-Trichloroethane	0.228	0.211	7.5	94	0.00
56 T	Ethyl methacrylate	0.285	0.273	4.2	93	0.00
57 T	1,3-Dichloropropane	0.391	0.372	4.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.135	-3.1	97	0.00
59 T	2-Hexanone	0.117	0.111	5.1	93	0.00
60 T	Dibromochloromethane	0.270	0.263	2.6	96	0.00
61 T	1,2-Dibromoethane	0.215	0.200	7.0	93	0.00
62 S	4-Bromofluorobenzene	0.383	0.388	-1.3	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.382	0.379	0.8	98	0.00
65 PM	Chlorobenzene	1.027	1.014	1.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.359	0.358	0.3	98	0.00
67 C	Ethyl Benzene	1.853	1.862	-0.5#	98	0.00
68 T	m/p-Xylenes	0.716	0.714	0.3	97	0.00
69 T	o-Xylene	0.658	0.658	0.0	98	0.00
70 T	Styrene	1.130	1.126	0.4	95	0.00
71 P	Bromoform	0.188	0.181	3.7	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.591	3.590	0.0	99	0.00
74 T	N-amyl acetate	0.695	0.655	5.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.567	0.513	9.5	94	0.00
76 T	1,2,3-Trichloropropane	0.359	0.407	-13.4	119	0.00
77 T	Bromobenzene	0.858	0.821	4.3	95	0.00
78 T	n-propylbenzene	4.156	4.159	-0.1	98	0.00
79 T	2-Chlorotoluene	2.336	2.296	1.7	97	0.00
80 T	1,3,5-Trimethylbenzene	3.031	3.014	0.6	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.177	6.8	95	0.00
82 T	4-Chlorotoluene	2.452	2.409	1.8	96	0.00
83 T	tert-Butylbenzene	2.691	2.672	0.7	97	0.00
84 T	1,2,4-Trimethylbenzene	2.990	2.963	0.9	97	0.00
85 T	sec-Butylbenzene	3.679	3.680	-0.0	98	0.00
86 T	p-Isopropyltoluene	3.427	3.453	-0.8	99	0.00
87 T	1,3-Dichlorobenzene	1.662	1.610	3.1	96	0.00
88 T	1,4-Dichlorobenzene	1.651	1.594	3.5	97	0.00
89 T	n-Butylbenzene	3.094	3.135	-1.3	98	0.00

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102819\
Data File : VW013759.D
Acq On : 28 Oct 2019 11:43
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 29 10:59:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 23 14:11:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.596	0.594	0.3	97	0.00
91 T	1,2-Dichlorobenzene	1.441	1.362	5.5	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.086	0.077	10.5	94	0.00
93 T	1,2,4-Trichlorobenzene	1.049	1.036	1.2	97	0.00
94 T	Hexachlorobutadiene	0.708	0.688	2.8	98	0.00
95 T	Naphthalene	1.644	1.582	3.8	92	0.00
96 T	1,2,3-Trichlorobenzene	0.904	0.883	2.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6