

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102918\
 Data File : VW006436.D
 Acq On : 27 Oct 2018 06:59
 Operator : SY/AP
 Sample : J5636-04RE
 Misc : 4.94G/5ML/MSVOA W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HP-GRID-DRE

Quant Time: Oct 29 04:11:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	177009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	245976	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	138480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	31043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	86660	59.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.10%	
35) Dibromofluoromethane	7.88	113	91913	63.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.48%	
50) Toluene-d8	10.33	98	261347	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
62) 4-Bromofluorobenzene	12.62	95	46148	22.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.44%	
Target Compounds						
20) Methylene Chloride	4.91	84	6617	2.722	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102918\
Data File : VW006436.D
Acq On : 27 Oct 2018 06:59
Operator : SY/AP
Sample : J5636-04RE
Misc : 4.94G/5ML/MSVOA W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HP-GRID-DRE

Quant Time: Oct 29 04:11:43 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

