

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102918\
 Data File : VW006422.D
 Acq On : 27 Oct 2018 00:06
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 04:54:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.247	0.271	-9.7	76	0.00
3 P	Chloromethane	0.318	0.351	-10.4	83	0.00
4 C	Vinyl Chloride	0.455	0.505	-11.0#	80	0.00
5 T	Bromomethane	0.414	0.466	-12.6	85	0.00
6 T	Chloroethane	0.326	0.391	-19.9	87	0.00
7 T	Trichlorofluoromethane	0.424	0.490	-15.6	82	0.00
8 T	Diethyl Ether	0.225	0.246	-9.3	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.444	0.507	-14.2	85	0.00
10 T	Methyl Iodide	0.739	0.848	-14.7	84	0.00
11 T	Tert butyl alcohol	0.028	0.028	0.0	79	0.00
12 CM	1,1-Dichloroethene	0.418	0.471	-12.7#	85	0.00
13 T	Acrolein	0.027	0.024	11.1	70	0.00
14 T	Allyl chloride	0.597	0.651	-9.0	81	0.00
15 T	Acrylonitrile	0.081	0.096	-18.5	89	0.00
16 T	Acetone	0.089	0.087	2.2	89	0.00
17 T	Carbon Disulfide	1.261	1.274	-1.0	74	0.00
18 T	Methyl Acetate	0.209	0.237	-13.4	84	0.00
19 T	Methyl tert-butyl Ether	0.636	0.717	-12.7	84	0.00
20 T	Methylene Chloride	0.484	0.528	-9.1	90	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.547	-17.4	87	0.00
22 T	Diisopropyl ether	1.180	1.387	-17.5	87	0.00
23 T	Vinyl Acetate	0.697	0.792	-13.6	85	0.00
24 P	1,1-Dichloroethane	0.762	0.918	-20.5#	89	0.00
25 T	2-Butanone	0.112	0.127	-13.4	88	0.00
26 T	2,2-Dichloropropane	0.537	0.544	-1.3	78	0.00
27 T	cis-1,2-Dichloroethene	0.515	0.631	-22.5#	91	0.00
28 T	Bromochloromethane	0.294	0.342	-16.3	80	0.00
29 T	Tetrahydrofuran	0.069	0.077	-11.6	85	0.00
30 C	Chloroform	0.830	1.018	-22.7#	91	0.00
31 T	Cyclohexane	0.759	0.762	-0.4	79	0.00
32 T	1,1,1-Trichloroethane	0.749	0.845	-12.8	84	0.00
33 S	1,2-Dichloroethane-d4	0.415	0.425	-2.4	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.295	0.308	-4.4	75	0.00
36 T	1,1-Dichloropropene	0.461	0.531	-15.2	86	0.00
37 T	Ethyl Acetate	0.171	0.193	-12.9	89	0.00
38 T	Carbon Tetrachloride	0.505	0.563	-11.5	83	0.00
39 T	Methylcyclohexane	0.613	0.674	-10.0	83	0.00
40 TM	Benzene	1.264	1.538	-21.7#	91	0.00
41 T	Methacrylonitrile	0.103	0.112	-8.7	80	0.00
42 TM	1,2-Dichloroethane	0.371	0.447	-20.5#	90	0.00
43 T	Isopropyl Acetate	0.350	0.383	-9.4	83	0.00
44 TM	Trichloroethene	0.388	0.468	-20.6#	90	0.00
45 C	1,2-Dichloropropane	0.303	0.366	-20.8#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102918\
 Data File : VW006422.D
 Acq On : 27 Oct 2018 00:06
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 04:54:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.172	0.213	-23.8#	92	0.00
47 T	Bromodichloromethane	0.418	0.496	-18.7	88	0.00
48 T	Methyl methacrylate	0.166	0.183	-10.2	82	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	88	0.00
50 S	Toluene-d8	1.124	1.132	-0.7	73	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.198	-13.1	86	0.00
52 CM	Toluene	0.848	1.020	-20.3#	90	0.00
53 T	t-1,3-Dichloropropene	0.416	0.475	-14.2	84	0.00
54 T	cis-1,3-Dichloropropene	0.478	0.557	-16.5	86	0.00
55 T	1,1,2-Trichloroethane	0.241	0.300	-24.5#	93	0.00
56 T	Ethyl methacrylate	0.296	0.349	-17.9	88	0.00
57 T	1,3-Dichloropropane	0.399	0.488	-22.3#	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.113	0.124	-9.7	85	0.00
59 T	2-Hexanone	0.125	0.136	-8.8	84	0.00
60 T	Dibromochloromethane	0.301	0.359	-19.3	87	0.00
61 T	1,2-Dibromoethane	0.241	0.298	-23.7#	91	0.00
62 S	4-Bromofluorobenzene	0.422	0.438	-3.8	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	73	0.00
64 T	Tetrachloroethene	0.388	0.464	-19.6	92	0.00
65 PM	Chlorobenzene	1.070	1.288	-20.4#	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.452	-18.9	88	0.00
67 C	Ethyl Benzene	1.855	2.187	-17.9#	89	0.00
68 T	m/p-Xylenes	0.733	0.873	-19.1	90	0.00
69 T	o-Xylene	0.699	0.837	-19.7	90	0.00
70 T	Styrene	1.136	1.385	-21.9#	91	0.00
71 P	Bromoform	0.212	0.239	-12.7	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.532	4.013	-13.6	89	0.00
74 T	N-amyl acetate	0.686	0.719	-4.8	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.537	0.637	-18.6	91	0.00
76 T	1,2,3-Trichloropropane	0.382	0.484	-26.7#	102	0.00
77 T	Bromobenzene	0.865	1.023	-18.3	92	0.00
78 T	n-propylbenzene	4.110	4.678	-13.8	88	0.00
79 T	2-Chlorotoluene	2.309	2.655	-15.0	88	0.00
80 T	1,3,5-Trimethylbenzene	2.978	3.433	-15.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.160	0.166	-3.8	79	0.00
82 T	4-Chlorotoluene	2.419	2.819	-16.5	90	0.00
83 T	tert-Butylbenzene	2.664	3.045	-14.3	89	0.00
84 T	1,2,4-Trimethylbenzene	3.013	3.518	-16.8	90	0.00
85 T	sec-Butylbenzene	3.764	4.301	-14.3	88	0.00
86 T	p-Isopropyltoluene	3.404	3.931	-15.5	88	0.00
87 T	1,3-Dichlorobenzene	1.722	2.042	-18.6	92	0.00
88 T	1,4-Dichlorobenzene	1.714	2.023	-18.0	92	0.00
89 T	n-Butylbenzene	3.114	3.549	-14.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW102918\
Data File : VW006422.D
Acq On : 27 Oct 2018 00:06
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 04:54:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.639	-10.7	82	0.00
91 T	1,2-Dichlorobenzene	1.536	1.835	-19.5	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.098	-2.1	79	0.00
93 T	1,2,4-Trichlorobenzene	1.182	1.352	-14.4	91	0.00
94 T	Hexachlorobutadiene	0.742	0.833	-12.3	88	0.00
95 T	Naphthalene	1.966	2.219	-12.9	89	0.00
96 T	1,2,3-Trichlorobenzene	1.022	1.176	-15.1	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6