

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102919\
 Data File : VW013792.D
 Acq On : 29 Oct 2019 14:42
 Operator : SY/VA
 Sample : K5535-09
 Misc : 5.10G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BEK31

Quant Time: Oct 30 06:34:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 30 06:58:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	351410	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	320438	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	137314	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	105744	24.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.12%
7) Chloroethane-d5	2.89	69	85792	21.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.28%
10) 1,1-Dichloroethene-d2	4.03	63	159518	16.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.00%
20) 2-Butanone-d5	7.07	46	65633	42.56	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.12%
24) Chloroform-d	7.65	84	221083	22.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.92%
26) 1,2-Dichloroethane-d4	8.31	65	119065	23.16	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.64%
29) Benzene-d6	8.27	84	471645	24.50	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.00%
33) 1,2-Dichloropropane-d6	9.27	67	141081	24.54	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.16%
37) Toluene-d8	10.32	98	424485	23.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.64%
38) trans-1,3-Dichloropropene-	10.58	79	57658	22.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.92%
39) 2-Hexanone-d5	10.92	63	52248	47.53	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.06%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	107101	22.74	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.96%
61) 1,2-Dichlorobenzene-d4	13.85	152	110166	22.70	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.80%

Target Compounds

					Ovalue
13) Acetone	4.12	43	38684	22.602	ug/L 97
16) Methylene chloride	4.93	84	13336	1.982	ug/L 83
53) Ethylbenzene	11.73	91	41806	1.532	ug/L 98
54) m,p-Xylene	11.83	106	85756	8.131	ug/L 97
55) o-xylene	12.16	106	30189	3.016	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102919\
Data File : VW013792.D
Acq On : 29 Oct 2019 14:42
Operator : SY/VA
Sample : K5535-09
Misc : 5.10G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BEK31

Quant Time: Oct 30 06:34:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 30 06:58:53 2019
Response via : Initial Calibration

