

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW102919\
 Data File : VW013805.D
 Acq On : 29 Oct 2019 21:17
 Operator : SY/VA
 Sample : K5549-20
 Misc : 5.55G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAQ73

Quant Time: Oct 30 07:25:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 30 08:16:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	296701	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	289994	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	161818	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	97667	26.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.32%
7) Chloroethane-d5	2.88	69	79173	23.30	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.20%
10) 1,1-Dichloroethene-d2	4.02	63	147511	18.35	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.40%
20) 2-Butanone-d5	7.07	46	87581	67.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.54%
24) Chloroform-d	7.65	84	207093	24.94	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.76%
26) 1,2-Dichloroethane-d4	8.31	65	118044	27.19	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.76%
29) Benzene-d6	8.27	84	441836	25.36	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.44%
33) 1,2-Dichloropropane-d6	9.27	67	135853	26.11	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.44%
37) Toluene-d8	10.32	98	405368	25.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.92%
38) trans-1,3-Dichloropropene-	10.58	79	59703	25.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.72%
39) 2-Hexanone-d5	10.92	63	72349	72.73	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	145.46%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	131643	30.89	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	123.56%#
61) 1,2-Dichlorobenzene-d4	13.85	152	151030	26.41	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.64%

Target Compounds					Ovalue
13) Acetone	4.12	43	13671	9.461 ug/L	65
15) Methyl Acetate	4.67	43	48640	20.600 ug/L	99
44) Toluene	10.38	91	13911	0.622 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW102919\
Data File : VW013805.D
Acq On : 29 Oct 2019 21:17
Operator : SY/VA
Sample : K5549-20
Misc : 5.55G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAQ73

Quant Time: Oct 30 07:25:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 30 08:16:17 2019
Response via : Initial Calibration

