

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW103019\
 Data File : VW013843.D
 Acq On : 30 Oct 2019 17:45
 Operator : SY/VA
 Sample : K5596-05
 Misc : 3.78G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAQA6

Quant Time: Oct 31 07:21:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 31 05:29:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	306356	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	290507	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	131610	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	86162	22.92	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.68%
7) Chloroethane-d5	2.89	69	73302	20.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.56%
10) 1,1-Dichloroethene-d2	4.02	63	127756	15.39	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.56%
20) 2-Butanone-d5	7.07	46	74877	55.70	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.40%
24) Chloroform-d	7.65	84	185284	21.61	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.44%
26) 1,2-Dichloroethane-d4	8.30	65	109271	24.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.52%
29) Benzene-d6	8.27	84	386758	22.16	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.64%
33) 1,2-Dichloropropane-d6	9.27	67	123104	23.62	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.48%
37) Toluene-d8	10.32	98	342756	21.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.16%
38) trans-1,3-Dichloropropene-	10.58	79	50208	21.35	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.40%
39) 2-Hexanone-d5	10.93	63	61102	61.32	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.64%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	114895	26.91	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.64%
61) 1,2-Dichlorobenzene-d4	13.85	152	106520	22.90	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.60%

Target Compounds					Qvalue
13) Acetone	4.12	43	11655	7.811 ug/L	95
14) Carbon disulfide	4.38	76	6114	0.439 ug/L	98
15) Methyl Acetate	4.67	43	3497	1.434 ug/L #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW103019\
Data File : VW013843.D
Acq On : 30 Oct 2019 17:45
Operator : SY/VA
Sample : K5596-05
Misc : 3.78G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAQA6

Quant Time: Oct 31 07:21:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101819S.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 31 05:29:18 2019
Response via : Initial Calibration

