

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW103020\
 Data File : VW017032.D
 Acq On : 30 Oct 2020 10:41
 Operator : SY/VA
 Sample : VW1030SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK09

Quant Time: Oct 31 05:00:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 31 05:42:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	467178	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	421378	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	195708	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	124716	23.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.52%
7) Chloroethane-d5	2.90	69	107812	25.81	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.24%
10) 1,1-Dichloroethene-d2	4.03	63	193190	18.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.68%
20) 2-Butanone-d5	7.09	46	51877	51.08	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.16%
24) Chloroform-d	7.65	84	285953	26.54	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.16%
26) 1,2-Dichloroethane-d4	8.31	65	133634	26.35	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.40%
29) Benzene-d6	8.28	84	588021	27.17	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.68%
33) 1,2-Dichloropropane-d6	9.28	67	160672	27.50	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	110.00%
37) Toluene-d8	10.33	98	546428	26.81	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.24%
38) trans-1,3-Dichloropropene-	10.59	79	61948	24.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.92%
39) 2-Hexanone-d5	10.93	63	42139	51.37	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.74%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	115834	26.03	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.12%
61) 1,2-Dichlorobenzene-d4	13.85	152	194808	28.32	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	113.28%


Target Compounds					Ovalue
16) Methylene chloride	4.92	84	8058	1.305 ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW103020\
Data File : VW017032.D
Acq On : 30 Oct 2020 10:41
Operator : SY/VA
Sample : VW1030SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK09

Quant Time: Oct 31 05:00:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 31 05:42:11 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 1.305 ug/L
RT: 4.92 min Scan# 495
Delta R.T. 0.00 min
Lab File: VW017032.D
Acq: 30 Oct 2020 10:41

Instrument :
MSVOA_W
ClientSampleId :
VBLK09

Tot Ion: 84 Resp: 8058

Ion	Ratio	Lower	Upper
84	100		
49	100.1	73.8	137.0
86	71.7	47.4	88.0

