

Data File : VW006472.D
Acq On : 30 Oct 2018 01:00
Operator : SY/AP
Sample : J5624-08RE
Misc : 4.53G/5ML/MSVOA_W/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PA-SB-02BRE

Quant Time: Nov 04 01:28:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	150185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	216438	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	128688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	24192	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	76474	61.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.84%	
35) Dibromofluoromethane	7.88	113	80240	62.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.48%	
50) Toluene-d8	10.33	98	214270	44.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.08%	
62) 4-Bromofluorobenzene	12.62	95	35706	19.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.08%	
Target Compounds						
20) Methylene Chloride	4.92	84	5187	2.387	ug/l #	Qvalue 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW006472.D
Acq On : 30 Oct 2018 01:00
Operator : SY/AP
Sample : J5624-08RE
Misc : 4.53G/5ML/MSVOA_W/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PA-SB-02BRE

Quant Time: Nov 04 01:28:04 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

