

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW103118\
 Data File : VW006475.D
 Acq On : 30 Oct 2018 02:18
 Operator : SY/AP
 Sample : J5625-04RE
 Misc : 4.84G/5ML/MSVOA W/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-02RE

Quant Time: Oct 30 07:23:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	195040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	273028	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	195550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	60023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	93021	57.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.06%	
35) Dibromofluoromethane	7.89	113	95420	59.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.30%	
50) Toluene-d8	10.33	98	293516	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	12.62	95	71069	30.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.66%	
Target Compounds						
20) Methylene Chloride	4.91	84	5196	1.457	ug/l #	89
68) m/p-Xylenes	11.84	106	6174	2.153	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW103118\
Data File : VW006475.D
Acq On : 30 Oct 2018 02:18
Operator : SY/AP
Sample : J5625-04RE
Misc : 4.84G/5ML/MSVOA W/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-02RE

Quant Time: Oct 30 07:23:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

