

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW103118\
 Data File : VW006476.D
 Acq On : 30 Oct 2018 02:44
 Operator : SY/AP
 Sample : J5625-06RE
 Misc : 4.87G/5ML/MSVOA W/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-03RE

Quant Time: Oct 30 07:25:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	209026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	286539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	191294	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	52405	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	98505	56.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.68%	
35) Dibromofluoromethane	7.88	113	102464	60.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.04%	
50) Toluene-d8	10.33	98	294997	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
62) 4-Bromofluorobenzene	12.62	95	67423	27.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.74%	
Target Compounds						
8) Diethyl Ether	3.68	74	1083	1.150	ug/l #	53
68) m/p-Xylenes	11.84	106	2869	1.023	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW103118\
Data File : VW006476.D
Acq On : 30 Oct 2018 02:44
Operator : SY/AP
Sample : J5625-06RE
Misc : 4.87G/5ML/MSVOA W/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-03RE

Quant Time: Oct 30 07:25:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

