

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW103118\
 Data File : VW006479.D
 Acq On : 30 Oct 2018 04:01
 Operator : SY/AP
 Sample : J5625-12RE
 Misc : 4.52G/5ML/MSVOA W/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-06RE

Quant Time: Oct 30 07:31:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	196967	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	271480	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	174697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	45537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	93966	57.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.08%	
35) Dibromofluoromethane	7.88	113	97503	60.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.56%	
50) Toluene-d8	10.33	98	266997	43.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.50%	
62) 4-Bromofluorobenzene	12.62	95	59149	25.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	51.60%	
Target Compounds						
20) Methylene Chloride	4.92	84	7331	2.703	ug/l	94
68) m/p-Xylenes	11.84	106	4250	1.659	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW103118\
Data File : VW006479.D
Acq On : 30 Oct 2018 04:01
Operator : SY/AP
Sample : J5625-12RE
Misc : 4.52G/5ML/MSVOA W/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-06RE

Quant Time: Oct 30 07:31:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

