

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW103118\
 Data File : VW006482.D
 Acq On : 30 Oct 2018 05:19
 Operator : SY/AP
 Sample : J5205-05
 Misc : 6.17G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TR-05-102518

Quant Time: Oct 30 07:43:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	213848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	313359	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	255130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	112452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	106101	59.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.70%	
35) Dibromofluoromethane	7.89	113	107252	57.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.84%	
50) Toluene-d8	10.33	98	333732	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
62) 4-Bromofluorobenzene	12.62	95	111084	41.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.96%	
Target Compounds						
16) Acetone	4.15	43	12567	24.425	ug/l	99
18) Methyl Acetate	4.68	43	2127	2.384	ug/l	99
20) Methylene Chloride	4.91	84	8595	3.053	ug/l	94
52) Toluene	10.39	92	5500	1.035	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW103118\
Data File : VW006482.D
Acq On : 30 Oct 2018 05:19
Operator : SY/AP
Sample : J5205-05
Misc : 6.17G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TR-05-102518

Quant Time: Oct 30 07:43:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

