

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110118\
 Data File : VW006523.D
 Acq On : 31 Oct 2018 14:01
 Operator : SY/AP
 Sample : J5711-01
 Misc : 5.84G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1

Quant Time: Oct 31 15:43:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	345356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	496989	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	457506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	266862	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	116566	40.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.42%	
35) Dibromofluoromethane	7.89	113	121188	41.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.54%	
50) Toluene-d8	10.33	98	469868	42.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.12%	
62) 4-Bromofluorobenzene	12.62	95	178656	42.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.14%	
Target Compounds						
					Qvalue	
16) Acetone	4.15	43	9461	2.986	ug/l	91
20) Methylene Chloride	4.92	84	22873	6.122	ug/l	92
27) cis-1,2-Dichloroethene	7.17	96	12232	3.441	ug/l	91
39) Methylcyclohexane	9.34	83	31891	5.237	ug/l	97
73) Isopropylbenzene	12.47	105	35330	1.874	ug/l	98
78) n-propylbenzene	12.81	91	147957	6.746	ug/l	100
85) sec-Butylbenzene	13.39	105	105985	5.276	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110118\
Data File : VW006523.D
Acq On : 31 Oct 2018 14:01
Operator : SY/AP
Sample : J5711-01
Misc : 5.84G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1

Quant Time: Oct 31 15:43:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

