

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110118\
 Data File : VW006530.D
 Acq On : 31 Oct 2018 17:02
 Operator : SY/AP
 Sample : J5711-05
 Misc : 6.02G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-6

Quant Time: Oct 31 17:31:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	382726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	548783	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	483600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	224088	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	129987	40.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.94%	
35) Dibromofluoromethane	7.88	113	132503	40.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.72%	
50) Toluene-d8	10.33	98	506471	41.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.10%	
62) 4-Bromofluorobenzene	12.62	95	165353	35.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.38%	
Target Compounds						
20) Methylene Chloride	4.91	84	24427	5.838	ug/l	90
64) Tetrachloroethene	10.87	164	36781	9.797	ug/l	98
95) Naphthalene	15.38	128	25363	2.879	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110118\
Data File : VW006530.D
Acq On : 31 Oct 2018 17:02
Operator : SY/AP
Sample : J5711-05
Misc : 6.02G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-6

Quant Time: Oct 31 17:31:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

