

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110220\
 Data File : VW017041.D
 Acq On : 02 Nov 2020 12:38
 Operator : SY/VA
 Sample : L4543-14
 Misc : 6.20G/10ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG472

Quant Time: Nov 03 04:56:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 03 04:48:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	550408	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	493690	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	228515	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	140053	22.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.20%
7) Chloroethane-d5	2.90	69	115554	23.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.92%
10) 1,1-Dichloroethene-d2	4.03	63	206931	16.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.00%
20) 2-Butanone-d5	7.08	46	47903	40.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.06%
24) Chloroform-d	7.65	84	286115	22.54	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.16%
26) 1,2-Dichloroethane-d4	8.31	65	130660	21.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.48%
29) Benzene-d6	8.28	84	594011	23.42	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.68%
33) 1,2-Dichloropropane-d6	9.28	67	162157	23.69	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.76%
37) Toluene-d8	10.33	98	558916	23.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.64%
38) trans-1,3-Dichloropropene-	10.58	79	62619	21.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.20%
39) 2-Hexanone-d5	10.93	63	39186	40.77	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.54%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	107062	20.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.12%
61) 1,2-Dichlorobenzene-d4	13.86	152	194007	24.16	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.64%

Target Compounds


					Ovalue
13) Acetone	4.13	43	4518	5.611 ug/L	72
14) Carbon disulfide	4.40	76	7831	0.359 ug/L	98
16) Methylene chloride	4.92	84	20165	2.771 ug/L	98
40) 1,2-Dichloropropane	9.28	63	17259	2.834 ug/L #	84
42) cis-1,3-Dichloropropene	10.04	75	4407	0.463 ug/L #	79
43) 4-Methyl-2-pentanone	10.21	43	2223	0.770 ug/L #	85
45) trans-1,3-Dichloropropene	10.58	75	2409	0.305 ug/L #	76
59) 1,2,3-Trichloropropane	13.56	75	23843	6.367 ug/L	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110220\
Data File : VW017041.D
Acq On : 02 Nov 2020 12:38
Operator : SY/VA
Sample : L4543-14
Misc : 6.20G/10ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG472

Quant Time: Nov 03 04:56:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 03 04:48:47 2020
Response via : Initial Calibration





#13

Acetone

Concen: 5.611 ug/L

RT: 4.13 min Scan# 366

Delta R.T. -0.00 min

Lab File: VW017041.D

Acq: 02 Nov 2020 12:38

Instrument :

MSVOA_W

ClientSampleId :


BG472

Tot Ion: 43 Resp: 4518

Ion Ratio Lower Upper

43 100

58 18.5 0.0 69.8



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

4.13



1500

1000

500

0

Time--> 4.05 4.10 4.15 4.20



#14

Carbon disulfide

Concen: 0.359 ug/L

RT: 4.40 min Scan# 409

Delta R.T. 0.01 min

Lab File: VW017041.D


Acq: 02 Nov 2020 12:38

Tgt Ion: 76 Resp: 7831

Ion Ratio Lower Upper

76 100

78 8.2 7.3 10.9



Abundance Ion 76.00 (75.70 to 76.70): VW

Ion 78.00 (77.70 to 78.70): VW

4.40

3000

2500

2000


1500


1000

500

0

Time--> 4.30 4.35 4.40 4.45 4.50









#16
Methylene chloride
Concen: 2.771 ug/L
RT: 4.92 min Scan# 495
Delta R.T. -0.00 min
Lab File: VW017041.D
Acq: 02 Nov 2020 12:38




Instrument :
MSVOA_W
ClientSampled :
BG472


Tot Ion: 84 Resp: 20165
Ion Ratio Lower Upper
84 100
49 104.7 73.8 137.0
86 65.5 47.4 88.0



#40
1,2-Dichloropropane
Concen: 2.834 ug/L
RT: 9.28 min Scan# 1210
Delta R.T. -0.09 min
Lab File: VW017041.D
Acq: 02 Nov 2020 12:38

Tgt Ion: 63 Resp: 17259
Ion Ratio Lower Upper
63 100
112 0.3 4.5 6.7#





#42

cis-1,3-Dichloropropene

Concen: 0.463 ug/L

RT: 10.04 min Scan# 1335

Delta R.T. -0.03 min

Lab File: VW017041.D

Acq: 02 Nov 2020 12:38

Instrument :

MSVOA_W

ClientSampled :





BG472

Tot Ion: 75 Resp: 4407

Ion Ratio Lower Upper

75 100

77 44.3 22.7 42.1#



#43

4-Methyl-2-pentanone

Concen: 0.770 ug/L

RT: 10.21 min Scan# 1363

Delta R.T. -0.00 min

Lab File: VW017041.D

Acq: 02 Nov 2020 12:38




Tgt Ion: 43 Resp: 2223


Ion Ratio Lower Upper

43 100

58 32.4 34.4 51.6#

100 14.7 16.2 24.2#





#45

trans-1,3-Dichloropropene

Concen: 0.305 ug/L

RT: 10.58 min Scan# 1423

Delta R.T. -0.03 min

Lab File: VW017041.D

Acq: 02 Nov 2020 12:38

Instrument :

MSVOA_W

ClientSampled :





BG472

Tot Ion: 75 Resp: 2409

Ion Ratio Lower Upper

75 100

77 45.8 22.7 42.3#



#59

1,2,3-Trichloropropane

Concen: 6.367 ug/L

RT: 13.56 min Scan# 1912

Delta R.T. 0.79 min

Lab File: VW017041.D

Acq: 02 Nov 2020 12:38

Tgt Ion: 75 Resp: 23843

Ion Ratio Lower Upper

75 100

77 37.1 25.8 38.8

