

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110220\
 Data File : VW017041.D
 Acq On : 02 Nov 2020 12:38
 Operator : SY/VA
 Sample : L4543-14
 Misc : 6.20G/10ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG472

Quant Time: Nov 03 05:31:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 03 04:48:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	550408	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	493690	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	228515	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	140053	22.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.20%
7) Chloroethane-d5	2.90	69	115554	23.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.92%
10) 1,1-Dichloroethene-d2	4.03	63	206931	16.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.00%
20) 2-Butanone-d5	7.08	46	47903	40.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.06%
24) Chloroform-d	7.65	84	286115	22.54	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.16%
26) 1,2-Dichloroethane-d4	8.31	65	130660	21.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.48%
29) Benzene-d6	8.28	84	594011	23.42	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.68%
33) 1,2-Dichloropropane-d6	9.28	67	162157	23.69	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.76%
37) Toluene-d8	10.33	98	558916	23.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.64%
38) trans-1,3-Dichloropropene-	10.58	79	62619	21.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.20%
39) 2-Hexanone-d5	10.93	63	39186	40.77	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.54%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	107062	20.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.12%
61) 1,2-Dichlorobenzene-d4	13.86	152	194007	24.16	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.64%

Target Compounds

					Ovalue
13) Acetone	4.13	43	4518	5.611 ug/L	72
14) Carbon disulfide	4.40	76	7831	0.359 ug/L	98
16) Methylene chloride	4.92	84	20165	2.771 ug/L	98
43) 4-Methyl-2-pentanone	10.21	43	2223	0.770 ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110220\
Data File : VW017041.D
Acq On : 02 Nov 2020 12:38
Operator : SY/VA
Sample : L4543-14
Misc : 6.20G/10ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG472

Quant Time: Nov 03 05:31:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 03 04:48:47 2020
Response via : Initial Calibration

