

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110220\
 Data File : VW017043.D
 Acq On : 02 Nov 2020 13:24
 Operator : SY/VA
 Sample : L4543-16
 Misc : 4.98G/10ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG474

Quant Time: Nov 03 05:38:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 03 04:48:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	520804	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	462835	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	215968	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	129496	21.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.20%
7) Chloroethane-d5	2.90	69	108551	23.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.24%
10) 1,1-Dichloroethene-d2	4.03	63	196378	16.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.20%
20) 2-Butanone-d5	7.08	46	44046	38.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.80%
24) Chloroform-d	7.65	84	274330	22.84	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.36%
26) 1,2-Dichloroethane-d4	8.31	65	124846	22.08	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.32%
29) Benzene-d6	8.28	84	565618	23.79	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.16%
33) 1,2-Dichloropropane-d6	9.28	67	152620	23.78	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.12%
37) Toluene-d8	10.33	98	531699	23.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.00%
38) trans-1,3-Dichloropropene-	10.58	79	58543	21.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.96%
39) 2-Hexanone-d5	10.93	63	37346	41.45	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.90%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	100567	20.57	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.28%
61) 1,2-Dichlorobenzene-d4	13.86	152	179449	23.64	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.56%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	22395	3.252 ug/L	93
35) Trichloroethene	9.10	95	2551	0.360 ug/L	94
47) Tetrachloroethene	10.87	164	1229114	188.692 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110220\
Data File : VW017043.D
Acq On : 02 Nov 2020 13:24
Operator : SY/VA
Sample : L4543-16
Misc : 4.98G/10ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG474

Quant Time: Nov 03 05:38:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 03 04:48:47 2020
Response via : Initial Calibration

