

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110220\
 Data File : VW017045.D
 Acq On : 02 Nov 2020 14:11
 Operator : SY/VA
 Sample : L4543-15
 Misc : 5.36G/10ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG473

Quant Time: Nov 03 05:46:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 03 04:48:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	564997	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	491050	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	228805	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	146684	22.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.00%
7) Chloroethane-d5	2.90	69	120593	23.87	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.48%
10) 1,1-Dichloroethene-d2	4.03	63	220947	17.42	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.68%
20) 2-Butanone-d5	7.08	46	44650	36.35	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.70%
24) Chloroform-d	7.65	84	307483	23.59	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.36%
26) 1,2-Dichloroethane-d4	8.31	65	135378	22.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.28%
29) Benzene-d6	8.28	84	643593	25.51	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.04%
33) 1,2-Dichloropropane-d6	9.27	67	171226	25.15	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.60%
37) Toluene-d8	10.33	98	603681	25.42	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	101.68%
38) trans-1,3-Dichloropropene-	10.59	79	61915	21.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.72%
39) 2-Hexanone-d5	10.93	63	37151	38.86	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.72%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	108473	20.91	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.64%
61) 1,2-Dichlorobenzene-d4	13.85	152	206718	25.71	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.84%

Target Compounds

					Ovalue
13) Acetone	4.14	43	2911	3.522 ug/L	96
14) Carbon disulfide	4.39	76	16600	0.740 ug/L	99
16) Methylene chloride	4.92	84	23328	3.123 ug/L	99
43) 4-Methyl-2-pentanone	10.21	43	2111	0.735 ug/L #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110220\
Data File : VW017045.D
Acq On : 02 Nov 2020 14:11
Operator : SY/VA
Sample : L4543-15
Misc : 5.36G/10ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG473

Quant Time: Nov 03 05:46:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 03 04:48:47 2020
Response via : Initial Calibration

