

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110221\
 Data File : VW020743.D
 Acq On : 02 Nov 2021 16:02
 Operator : SY/VA
 Sample : VW1102SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1102SBL01

Quant Time: Nov 03 02:20:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W110121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 02 04:55:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	28589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	49392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	45821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	23515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	14383	66.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.280%
35) Dibromofluoromethane	7.884	113	16113	51.331	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.660%
50) Toluene-d8	10.323	98	52144	47.696	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.400%
62) 4-Bromofluorobenzene	12.615	95	16592	43.072	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110221\
Data File : VW020743.D
Acq On : 02 Nov 2021 16:02
Operator : SY/VA
Sample : VW1102SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1102SBL01

Quant Time: Nov 03 02:20:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W110121S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 02 04:55:07 2021
Response via : Initial Calibration

