




Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110320\
Data File : VW017060.D
Acq On : 03 Nov 2020 15:33
Operator : SY/VA
Sample : L4615-04
Misc : 5.74G/10ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Nov 04 04:11:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 04 04:03:33 2020
Response via : Initial Calibration



TIC: VW017060.D

(12) 1,1-Dichloroethene (T)

4.038min (-0.006) 1.34ug/L

response 9330

Ion	Exp%	Act%
96.00	100	100
61.00	144.90	244.77#
63.00	112.20	874.73#
0.00	0.00	0.00