

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110320\
 Data File : VW017065.D
 Acq On : 03 Nov 2020 17:30
 Operator : SY/VA
 Sample : VIBLK01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK01

Quant Time: Nov 04 05:50:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 04 04:03:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	426131	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	406530	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	195398	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	93116	18.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.72%
7) Chloroethane-d5	2.89	69	82681	21.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.80%
10) 1,1-Dichloroethene-d2	4.03	63	145632	15.22	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.88%
20) 2-Butanone-d5	7.08	46	52640	56.82	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.64%
24) Chloroform-d	7.65	84	233425	23.75	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	8.31	65	116240	25.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.52%
29) Benzene-d6	8.27	84	464989	22.27	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.08%
33) 1,2-Dichloropropane-d6	9.28	67	133673	23.71	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.84%
37) Toluene-d8	10.33	98	431678	21.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.84%
38) trans-1,3-Dichloropropene-	10.58	79	53343	22.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.20%
39) 2-Hexanone-d5	10.93	63	43550	55.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.06%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	112211	26.13	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.52%
61) 1,2-Dichlorobenzene-d4	13.85	152	168266	24.50	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.00%

Target Compounds

					Ovalue
16) Methylene chloride	4.92	84	7888	1.400 ug/L	92
43) 4-Methyl-2-pentanone	10.21	43	3327	1.399 ug/L #	84
47) Tetrachloroethene	10.87	164	17619	3.079 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110320\
Data File : VW017065.D
Acq On : 03 Nov 2020 17:30
Operator : SY/VA
Sample : VIBLK01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK01

Quant Time: Nov 04 05:50:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM102220S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 04 04:03:33 2020
Response via : Initial Calibration

