


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW110418\
 Data File : VW006600.D
 Acq On : 02 Nov 2018 13:36
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Title : SW846 8260
 Last Update : Thu Nov 01 03:48:08 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	22197	PASS
75	95	30	60	48.8	62957	PASS
95	95	100	100	100.0	129106	PASS
96	95	5	9	6.9	8947	PASS
173	174	0.00	2	0.4	480	PASS
174	95	50	100	95.8	123698	PASS
175	174	5	9	7.5	9268	PASS
176	174	95	101	96.4	119213	PASS
177	176	5	9	6.8	8066	PASS