

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110518\
 Data File : VW006650.D
 Acq On : 05 Nov 2018 21:16
 Operator : SY/AP
 Sample : J5700-04RE
 Misc : 5.63G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SI-4RE

Quant Time: Nov 06 03:19:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 01 03:48:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-7.95
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-8.85
63) Chlorobenzene-d5	11.65	117	276	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.57	152	271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	10.33	98	46	0.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	0.00%	
62) 4-Bromofluorobenzene	12.62	95	111	0.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	0.00%	
Target Compounds						
					Qvalue	
67) Ethyl Benzene	11.74	91	97	9.992	ug/l #	42
68) m/p-Xylenes	11.85	106	53	13.665	ug/l #	1
70) Styrene	12.19	104	174	28.905	ug/l #	27
80) 1,3,5-Trimethylbenzene	13.11	105	40	2.667	ug/l #	27
84) 1,2,4-Trimethylbenzene	13.26	105	73	4.745	ug/l #	30
85) sec-Butylbenzene	13.26	105	73	3.897	ug/l #	55
95) Naphthalene	15.38	128	144	14.716	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110518\
Data File : VW006650.D
Acq On : 05 Nov 2018 21:16
Operator : SY/AP
Sample : J5700-04RE
Misc : 5.63G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SI-4RE

Quant Time: Nov 06 03:19:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 01 03:48:08 2018
Response via : Initial Calibration

